



Update: AMI Monitor (#01-05-0163)

Date: 6 November 2008

Re: Service Update for CASMED Model AMI Monitors

Dear Valued Customer,

The purpose of this letter is to provide you with updates to our Service Policy for the Model AMI Monitor. Please note this service update only applies to the original green AMI Monitor. Service will continue on the tan AMI + monitors.

Unfortunately, due to the obsolescence of several internal components, CASMED will no longer be able to provide service, repair, or preventative maintenance for the Model AMI monitor effective **December 1, 2008**.

Please feel free to contact our Service Department regarding any questions you may have at 800-581-7806.

Our Customer Service Department is available for help with information on exciting new products and can be reached at 800-580-7032 or 203-488-6056 or you may visit our website at www.casmed.com.

We apologize for any inconvenience this may cause you and look forward to continuing our monitoring relationship with you.

Sincerely,

Beth Schmidt
Product Manager
CAS Medical Systems, Inc.
44 E Industrial Road
Branford, CT 06405
800-227-4414 Toll Free
203-315-6946 Direct
eschmidt@casmed.com